

Fluke VPS510-V ScopeMeter® Compact Probe



Características principales

- Ideal for small electronics applications
- Compact size designed for lower voltage applications
- Double insulated, 300 V CAT III rated
- 10:1 voltage ratio and 500 MHz bandwidth
- Available in red, blue, grey, and green

Descripción general del producto: Fluke VPS510-V ScopeMeter® Compact Probe

Use the Fluke VPS510-V (Green) Compact Voltage Probe with the [Fluke 190 Series III](#) and Series II Portable Oscilloscopes or with the [MDA-500 Series Motor Drive Analyzers](#). These double insulated compact probes offer a 10:1 voltage ratio, and 500 MHz bandwidth making them ideal for small electronics applications. For industrial applications consider Fluke [VPS410-II-V Voltage Probes](#) or [VPS421 Series High Voltage Probes](#).

Especificaciones: Fluke VPS510-V ScopeMeter® Compact Probe

Aenuation	10:1
Bandwidth	500 MHz
Max. input voltage	CAT III 300 V
Max. voltage to ground	CAT III 300 V
Input Resistance	10 MΩ
Input Capacitance	11.5 ± 1 pF
Compensation range	10...25 pF
Cable Length	1.2 m (3.9 ft)
Safety	IEC / EN 61010-031 UL 61010-031 CSA C22.2 No. 1010.2.031
Max. Altitude	2000 m
Pollution Degree	2
Protection Class	II (double insulation)

*Specification of probe is valid only when used with Fluke ScopeMeter 190-xxx models and MDA-500 series.

Modelos



Fluke VPS510-V

ScopeMeter® Compact Probe for electrical applications

- (1) Green voltage probe
- (1) Black hook clip for probe tip
- (1) Black ground lead with mini alligator clip
- (1) Ground spring for probe tip (CAT II 30 V max.)
- (1) Insulation sleeve for probe tip
- (1) BNC to probe tip adapter
- (1) Adjustment tool

Fluke. *Manteniendo su mundo en marcha.*

Fluke Corporation

Everett, WA 98206 EE.UU.

Para obtener información adicional En EE. UU.

(800) 443-5853

En Europa/Medio Oriente/África

+31 (0)40 267 5100

En Canadá (800)-36-FLUKE

www.fluke.com

Latin America

Tel: +1 (425) 446-5500

www.fluke.com/es-uy

©2025 Fluke Corporation. Reservados todos los derechos. Impreso en los Países Bajos. Información sujeta a modificación sin previo aviso.

04/2025

No está permitido modificar este documento sin autorización por escrito de Fluke Corporation.