

Dati tecnici**4180-APRT | Aperture, 50 mm (2 in), 4180 or 4181****Descrizione generale del prodotto: 4180-APRT | Aperture, 50 mm (2 in), 4180 or 4181**

Many radiometers require the use of an aperture to reduce the effect of scatter. This is a simple-to-use, 2-inch non-water cooled aperture designed for use with the 4180 or 4181 Precision Infrared Calibrators.

Modelli



4180-APRT

Aperture, 50 mm (2 in), 4180 or 4181

Fluke. *Keeping your world up and running.®*

Fluke Italia S.r.l.
Viale Lombardia 218
20861 Brugherio (MB)
Tel: +39 02 3600 2000
E-mail: cs.it@fluke.com
www.fluke.it

©2025 Fluke Corporation. Tutti i diritti riservati.
Dati passibili di modifiche senza preavviso.
04/2025

**Non sono ammesse modifiche al presente
documento senza autorizzazione scritta da parte
di Fluke Corporation.**